


Certificate of Calibration



Certificate No: SPCCAL00263
Manufacturer: Candura Instruments
Model: SPC Pro
Serial No: SPCPRO00263
Date Calibrated: 15-Sep-09

The product identified above is hereby certified according to calibration procedures and quality inspection performed by Candura Instruments technician.

Calibration Meter: Fluke 6100A Power Standard
Model No: 6100A
Serial No: 917452981

240V - Tolerance +/- 1.2V

| Channel | 240V | 120V | 24V |
|---------|-------|-------|------|
| V1 | 240.2 | 120.1 | 24.0 |

240V, 20A Flex- Tolerance +/- 0.1A

| Channel | 20A | 10A | 2A |
|---------|------|------|-----|
| I1 | 20.0 | 10.0 | 2.0 |
| I2 | 20.0 | 10.0 | 2.0 |
| I3 | 20.0 | 10.0 | 2.0 |

240V, 200A Flex - Tolerance +/- 1.0A

| Channel | 200A | 100A | 25A |
|---------|-------|-------|------|
| I1 | 200.0 | 100.1 | 24.9 |
| I2 | 200.0 | 100.0 | 24.9 |
| I3 | 200.0 | 100.0 | 24.9 |

240V, 2000A Flex - Tolerance +/- 10A

| Channel | 2000A | 1000A | 250A |
|---------|--------|--------|-------|
| I1 | 2000.1 | 1001.1 | 249.6 |
| I2 | 2000.1 | 1000.6 | 249.4 |
| I3 | 2000.0 | 1000.7 | 249.5 |
| Angle 1 | 0.0 | | |
| Angle 2 | 0.0 | | |
| Angle 3 | 0.0 | | |

Calibrated by: Bill Brown